

ARTYKUŁ ORYGINALNY / ORIGINAL PAPER

Otrzymano/Submitted: 19.04.2017 • Zaakceptowano/Accepted: 31.05.2017

© Akademia Medycyny

Application and utility of the endotracheal stylet for direct laryngoscopy. A prospective, randomized, cross-over manikin-based simulation trial**Wojciech Wieczorek^{1,2}, Marek Dąbrowski^{2,3}, Agata Dąbrowska^{2,3}, Aleksandra Patrańska⁴, Marcin Madziara⁵, Jacek Smereka⁶, Łukasz Szarpak^{2,5}**¹ Department of Anaesthesiology, Intensive Care and Emergency Medicine in Zabrze, Medical University of Silesia in Katowice, Poland² Polish Society of Medical Simulation, Poland³ Department of Rescue and Disaster Medicine, Poznan University of Medical Sciences, Poland⁴ Journal Club, Poznan University of Medical Sciences, Poland⁵ Department of Emergency Medicine, Medical University of Warsaw, Poland⁶ Department of Emergency Medical Service, Wrocław Medical University, Wrocław, Poland**Abstract**

Background. One of the most important priorities in managing life threatening conditions is maintaining the patency of the airway. The aim of the study was to evaluate the efficacy and duration of tracheal intubation with and without intubation stylet in simulated damage to the cervical spine. **Material and methods.** The study was prospective, randomized, crossover, manikin trial, with 27 paramedics participants. Paramedics intubated paediatric simulator using Macintosh laryngoscopes with and without endotracheal stylet. Participants performed intubation in two scenarios: Scenario A – normal airway; Scenario B – neck immobilization using cervical collar. The study measured overall success rate, first intubation attempt success rate, and time to intubation. **Results.** 27 paramedics participated in this trial. In normal airway scenario first intubation success rate was 100% for endotracheal tube with and without endotracheal stylet. Median time to intubation using distinct methods were 22 [IQR; 19-27.5]s vs. 18.5 [IQR; 17-26]s (with and without stylet, respectively). During cervical collar immobilization scenario success of first intubation attempt varied and amounted to 18.5% without stylet, and 48.1% for tube with stylet ($p < 0.001$). The overall intubation success rate was also higher when the stylet was used (88.9%) compared with tube without stylet (51.8%; $p = 0.014$). Median intubation time was 35.5 [IQR; 32-48]s for stylet, and 43.5 [IQR; 38-53]s for intubation without stylet ($p < 0.001$). **Conclusion.** For normal airways, the use of an intubation tube with a stylet does not increase the effectiveness of endotracheal intubation; however, with the use of cervical collar immobilization, stylet intubation was associated with a shorter procedure time and higher intubation efficacy. *Anestezjologia i Ratownictwo 2017; 11: 131-137.*

Keywords: endotracheal intubation, paediatric, immobilization, stylet, simulation

Introduction

One of the most important priorities in managing life threatening conditions is maintaining the patency of the airway [1,2]. Despite many methods of airway clearing from simple to advanced, intubation is always the most trusted and effective method. It provides continuous airway patency, but does not relieve us of the obligation to monitor the correct position of the intubation tube. The intubation itself is usually a procedure lasting a dozen or so seconds, but much depends the efficiency of the intubator [3,4]. Unexplained activity can aggravate the appearance of undesirable effects. The main indication for intubation of the patient is the possibility of mechanical ventilation and prevention of aspiration of gastric contents. Post introduction of the intubation tube technique, there has been increasing success in maintenance of the patients airways in certain scenarios.

The most common complications include: soft tissue damage to the airways (larynx, back of the throat, nasal cavity), which is usually manifested by minor transient bleeding; pituitary edema of the larynx with temporary difficulty in breathing or speaking which requires additional treatment (inhalation, medication); permanent damage to vocal cords that manifest as hoarseness; and damage to the gingiva or teeth by the laryngoscope [5,6]. Usually, the cause of complications lies in the unusual construction of the oral cavity and upper respiratory tract, which makes the procedure technically difficult to perform.

In pre-hospital conditions and emergency department, intubation may prove to be a difficult operation because there is little familiarity with the patient at hand, as well as possible injury to the patient, more so in the case of the paediatric population. The fact that the infant's head should not be overextended which poses a challenge to see the glottis (black vertical triangular slit with white cords). The use of cervical collar in stabilization of the cervical vertebra makes it more difficult to complete the procedure. Regardless of whether the stabilization is manual or maintained by the cervical collar, it reduces movement and increases the risk of complications. These actions require staff to be prepared both before and during intubation. Efficient performing of advanced clearing does not only depend on the person who directly performs the intubation, but also on assistants who are involved in the patient's life-threatening condition.

The aim of the study was to evaluate the efficacy and duration of tracheal intubation with and without intubation stylet in simulated damage to the cervical spine.

Material and methods

Study Design

The study is a follow-up study by the authors investigating the most effective method of endotracheal tracheal intubation in a spinal cord injury [7-9]. This was a prospective, randomized, crossover, manikin trial conducted at the Department of Emergency Medicine at the Medical University of Warsaw. The study protocol was approved by the Institutional Review Board of the Polish Society of Disaster Medicine (Approval: 05.02.2017.IRB), and was conducted in March of 2017.

Participants

The participants for this trial were recruited paramedics participating in Airway Management courses organized by Polish Society of Disaster Medicine. All participants agreed for voluntarily participation in the study; written informed consent was obtained from all of the participants. A total of 27 paramedics participated in the study.

Study Interventions

Prior to the study, each participant completed a respiratory tract respiratory training course that consisted of endotracheal intubation with a Macintosh and Miller blade laryngoscope and supraglottic airway devices. The target study was held two months after their initial training. During the study, practicing with the devices prior to the recorded attempts was not permitted in order to minimize the bias, as well as to increase the simulation's reliability and predictability of actual field scenarios.

During the study, participants intubated using the Miller laryngoscope (HEINE Optotechnik, Munich, Germany) with a 5.0 mm internal diameter (ID) tracheal tube. In the study of intubation two general scenarios were examined, with and without an intubation guide.

Participants in the study performed intubation in the following two rescue scenarios:

- Scenario A: Normal airway scenario
- Scenario B: Normal airway with a correctly fitted standard cervical immobilization collar (Stifneck

Select, Laerdal, Stavanger, Norway) applied to the manikin's neck to prevent movement of the cervical spine (figure 1).

All intubation attempts were performed on a standardized pediatric manikin representing 6-year-old boy, with a regular airway (SimJunior, Laerdal, Stavanger, Norway).



Figure I. Cervical spine immobilization scenario using cervical collar

Each paramedic performed a maximum three intubation attempts using both intubation methods

in the two scenarios. The order of method used (with or without endotracheal stylet) was randomized with a ratio of 1:1 prior the collection of data using Research Randomizer Software (randomizer.org). The procedure for randomization is shown in figure 2.)

Outcomes

Only one investigator collected data on a standardized form. All intubation attempts were recorded with the GoPro HERO5 camcorder, and the timing was precisely confirmed by playback of the recordings. The primary endpoint was endotracheal intubation success rate. Overall success rate of intubation was defined as successful endotracheal intubation with the initial method, during maximum three attempts. Successful intubation attempt was defined as the correct placement of the endotracheal tube into the trachea; confirmed by both the flow of air in inflation using the resuscitator bag and the visualization of the tube by the investigator. An intubation failure was defined by an intubation requiring a time greater than 60s [10]. The endpoints were time to intubation, Cormack & Lehane graduation [11], and ease of intubation. Time to intubation, was defined from the time of picking up the




Figure II. Randomization flow chart

airway device to ending with the first effective manual ventilation of the manikin's lungs which was identified by the participant as the manikin's chest rising. The ease of intubation score was rated by the participant using a standard 100 point Likert Scale ranging ("1" being very easy to intubate, "100" being very difficult).

Statistical analysis

The occurrence of normal distribution was confirmed by the Kolmogorov–Smirnov test. All statistical tests were 2-sided. Nonparametric tests were used for the data that did not have a normal distribution. In order to compare the difference and determine the statistical significance in the time needed to achieve a sufficient glottis view and first successful ventilation, the Wilcoxon test for paired observations was used. The McNemar test was used to evaluate the differences in success of intubation. The Cormack&Lehane grade and VAS score were all evaluated using the Stuart–Maxwell test. Percentages were used for qualitative variables and median with interquartile range (IQR) for quantitative variables. A result was considered as significant with p-value less than 0.05. Analysis was performed using statistical software Statistica 13.1 (Statsoft, Tulsa, OK, USA).

Results

Twenty-seven paramedics (12 female, 44.4%) volunteered to participate in this trial. Median age was 26.5 [IQR; 24-33] years, and median work experience in Emergency Medical Services was 4.5 [IQR; 2-7] years.

During scenario A (normal airways) the effectiveness of the first attempt at intubation both with and without stylet was 100%. The median time of intubation in the case of the guide was shorter than in the case of intubation with no conduit without guide (18.5 [IQR; 17-26] vs. 22 [IQR; 19-27.5], respectively; Table I). Cervical collar immobilization scenario (B) was associated with a decrease in the effectiveness of intubation both with and without a guide. The efficacy of the first endotracheal tube with stylet was 48.1%, and for endotracheal tube without stylet the efficacy was 18.5% (Table II). The overall efficacy of the tracheal intubation was 88.9%, and was statistically significantly higher than for intubation without guide (51.8%; $p = 0.014$). Median time using distinct methods varied, amounting to 35.5 [IQR; 32-48]s vs. 43.5 [IQR; 38-53]s (with and without stylet, respectively; figure 3).

Participants in the study found intubation of the patient with an intubation tube equipped with a guide an easier tracheal intubation method when compared to a tube without a guide. This relationship was related to both intubation in normal airways ($p = 0.007$) as well as intubation under cervical vertebra ($p < 0.001$).




Figure III. Median time to intubation

Discussion

According to the results of this study, the overall success rates as well as the first intubation attempt success rate during the normal airway scenario were similar between the tracheal tube, with and without stylet. However, during intubation in cervical spine scenario (B), intubation with stylet improved the efficiency of intubation and reduce the intubation time when compared to endotracheal tube without stylet method.

There are two major risks involving the intubation of patients with cervical spine injuries [12]. The first risk is prolonged intubation with the risk of vomiting and aspiration, generally occurring in non-fasting patients. The second risk is cervical spine excursion (particularly within the functional region of occiput C3), or defensive cervical spine movement leading to additional cervical spine and neurological damage in sedated patients [13,14]. In addition, more than three tracheal endotracheal intubation efforts may result in soft tissue swelling and bleeding, this will significantly impede subsequent intubation efforts leading to a situation described by the Difficult Airway Society as "cannot intubate, cannot ventilate" [15]. Due to this complication, first attempt success rate is important because fewer intubation attempts ultimately affect

overall intubation success, with repeated intubation attempts associated with the risk of further injury, hypoxemia, and hemodynamic changes [16-18].

Protection of the airway patency in patients with suspected traumatic injury is one of the basic skills that medical staff should attend to [19]. Advanced trauma life support guidelines requires continuous cervical immobilization of the patient by a semi-rigid cervical collar [1]. The cervical collar should not be removed until the cervical injury is excluded by imaging studies in Emergency Department [20]. As indicated by many studies of the use of the neck flange, it reduces the effectiveness of the first intubation test and prolongs the duration of the procedure [7,9,21,22]. This dependency is also reflected in the results of this study in which the introduction of a neck collar resulted in a decrease in the efficacy of a first endotracheal tracheal intubation trial without a guide by 81.5%, and by 51.9% in the case of an intubation tube with a guide.

The study used a laryngoscope with a Macintosh blade. In the study by Passi et al. of infants and children < 2 yrs, optimal laryngeal views may be obtained by either Miller size 1 blade epiglottis or with the Miller or MAC blades lifting the tongue base [23]. Similarly, Verghese and Kurdu [24] demonstrated that, in children aged 1-24, the Miller and the Macintosh blades provide similar laryngoscopic views and intubating conditions. Therefore the study was performed using the Macintosh laryngoscope.

In a study by Hippard et al. [25] of thirty experienced paediatric anaesthesia practitioners tracheal intubation was performed in an infant model, stating that median intubation time in normal airway using Miller blade was 26 [IQR; 20-32]. In turn Szarpak et al. [26], a study comparing the efficacy of video rigid flexion laryngoscope and Miller laryngoscope, the efficacy of Miller laryngoscope without stylet was 96.9%, and intubation time of 23 [IQR, 19-26] was comparable to the results obtained in our study. The same authors in the difficult airway scenario demonstrated direct laryngoscopy efficacy at 44.7% at intubation 41 [IQR; 34-45] seconds. Rodríguez-Núñez et al. [27] in addition, demonstrated that intubation time using direct laryngoscopy with endotracheal tube without

stylet during ongoing chest compression was 28.2 (20.4-34.4) s.

The use of the guide beyond the obvious benefits of increasing the effectiveness of endotracheal intubation may also carry some complications [28-30]. In the study by Komasa et al. [31] comparing the impact of postoperative pharyngeal pain on postoperative pharyngeal pain or inferiority in patients undergoing elective surgery, indicated that a stylet use can increase the incidence of postoperative pharyngeal pain.

The strengths of our study include the randomization of the collection, as well as the participants, the paramedics group are often forced to perform endotracheal intubation as well as to protect the traumatic patient at the pre-hospital stage. Another strength lied in the use of a medical simulator allowing for the standardization of the difficulties resulting from endotracheal intubation and, secondly, allowing for cross-examination without harm to a potential patient. However, this led to a study limitations of realistic reliability in the use of a dummy rather than clinical cases. Another limitation of the study is the use of only Macintosh laryngoscope; in most ambulances in Poland, the Miller and Macintosh laryngoscopes are most commonly used.

Conclusions

The use of the stylet may be useful during intubation in the scenario of immobilization of the cervical spine, but does not affect the improvement of intubation efficiency under normal airways. Further studies are needed to confirm the results.

Conflict of interest

None

Correspondence address:

✉ Marcin Madzala

Department of Emergency Medicine,

Medical University of Warsaw

4, Lindleya Str., 02-005 Warsaw, Poland

☎ (+48 22) 502 15 25

✉ mmadzala262@gmail.com

References

1. Sperka J, Hanson SJ, Hoffmann RG, Dasgupta M, Meyer MT. The Effects of Pediatric Advanced Life Support Guidelines on Pediatric Trauma Airway Management. *Pediatr Emerg Care*. 2016 Aug;32(8):499-503. doi: 10.1097/PEC.0000000000000759.
2. Barak M, Bahouth H, Leiser Y, Abu El-Naaj I. Airway Management of the Patient with Maxillofacial Trauma: Review of the Literature and Suggested Clinical Approach. *Biomed Res Int*. 2015;2015:724032. doi: 10.1155/2015/724032.
3. Vivek B, Sripriya R, Mishra G, Ravishankar M, Parthasarathy S. Comparison of success of tracheal intubation using Macintosh laryngoscope-assisted Bonfils fiberscope and Truview video laryngoscope in simulated difficult airway. *J Anaesthesiol Clin Pharmacol*. 2017 Jan-Mar;33(1):107-111. doi: 10.4103/0970-9185.202198.
4. Stein C, Gerber L, Curtin D, Oberem N, Wells M. A Comparison of Three Maneuvers and Their Effect on Laryngoscopic View, Time to Intubate, and Intubation Outcome by Novice Intubators in a Simulated Airway. *Prehosp Disaster Med*. 2017 Apr 6:1-5. doi: 10.1017/S1049023X17006392.
5. Eiamcharoenwit J, Itthisompaiboon N, Limpawattana P, Siriussawakul A. The Performance of the Intubation Difficulty Scale among Obese Parturients Undergoing Cesarean Section. *Biomed Res Int*. 2017;2017:3075756. doi: 10.1155/2017/3075756.
6. van Esch BF, Stegeman I, Smit AL. Comparison of laryngeal mask airway vs tracheal intubation: a systematic review on airway complications. *J Clin Anesth*. 2017 Feb;36:142-150. doi: 10.1016/j.jclinane.2016.10.004.
7. Gawlowski P, Smereka J, Madziala M, Szarpak L, Frass M, Robak O. Comparison of the Macintosh laryngoscope and blind intubation via the iGEL for Intubation With C-spine immobilization: A Randomized, crossover, manikin trial. *Am J Emerg Med*. 2017 Mar;35(3):484-487. doi: 10.1016/j.ajem.2016.11.064.
8. Madziala M. Endotracheal intubation during manual inline cervical stabilization performed by nurses. *Am J Emerg Med*. 2016 Dec;34(12):2456-2457. doi: 10.1016/j.ajem.2016.09.030.
9. Smereka J, Ladny JR, Naylor A, Ruetzler K, Szarpak L. C-MAC compared with direct laryngoscopy for intubation in patients with cervical spine immobilization: A manikin trial. *Am J Emerg Med*. 2017 Mar 16. pii: S0735-6757(17)30196-1. doi: 10.1016/j.ajem.2017.03.030.
10. Kurowski A, Szarpak L, Truszewski Z, Czyzewski L. Can the ETView VivaSight SL Rival Conventional Intubation Using the Macintosh Laryngoscope During Adult Resuscitation by Novice Physicians? A Randomized Crossover Manikin Study. *Medicine (Baltimore)*. 2015 May;94(21):e850. doi: 10.1097/MD.0000000000000850.
11. Cormack RS, Lehane J. Difficult tracheal intubation in obstetrics. *Anaesthesia*. 1984 Nov;39(11):1105-11.
12. January AM, Zebracki K, Chlan KM, Vogel LC. Poor sleep in adults with pediatric-onset spinal cord injury: associations with pain, health, and activity. *J Spinal Cord Med*. 2017 Apr 10:1-7. doi: 10.1080/10790268.2017.1308109.
13. Thesleff T, Kataja A, Öhman J, Luoto TM. Head injuries and the risk of concurrent cervical spine fractures. *Acta Neurochir (Wien)*. 2017 May;159(5):907-914. doi: 10.1007/s00701-017-3133-0.
14. Chu MW, Soleimani T, Evans TA, Fernandez SI, Spera L, Klene C, Zarzaur BL, Tholpady SS. C-spine injury and mandibular fractures: lifesaver broken in two spots. *J Surg Res*. 2016 Dec;206(2):386-390. doi: 10.1016/j.jss.2016.08.019.
15. McNiece WL, Dierdorf SF. The pediatric airway. *Semin Pediatr Surg*. 2004 Aug;13(3):152-65.
16. Martin LD, Mhyre JM, Shanks AM, Tremper KK, Kheterpal S. 3,423 emergency tracheal intubations at a university hospital: airway outcomes and complications. *Anesthesiology*. 2011; 114(1):42-8. doi: 10.1097/ALN.0b013e318201c415.
17. Combes X, Jabre P, Jbeili C, Leroux B, Bastuji-Garin S, Margenet A, Adnet F, Dhonneur G. Prehospital standardization of medical airway management: incidence and risk factors of difficult airway. *Acad Emerg Med*. 2006; 13(8):828-34.
18. Mort TC. Emergency tracheal intubation: complications associated with repeated laryngoscopic attempts. *Anesth Analg*. 2004;99(2):607-13.
19. Magnone S, Allegri A, Belotti E, Castelli CC, Ceresoli M, Coccolini F, Manfredi R, Merli C, Palamara F, Piazzalunga D, Valetti TM, Ansaloni L. Impact of ATLS guidelines, trauma team introduction, and 24-hour mortality due to severe trauma in a busy, metropolitan Italian hospital: A case control study. *Ulus Travma Acil Cerrahi Derg*. 2016; 22(3):242-6. doi: 10.5505/tjtes.2015.19540.
20. Rao PJ, Phan K, Mobbs RJ, Wilson D, Ball J. Cervical spine immobilization in the elderly population. *J Spine Surg*. 2016 Mar;2(1):41-6. doi: 10.21037/jss.2016.02.02.
21. Shraavanalakshmi D, Bidkar PU, Narmadalakshmi K, Lata S, Mishra SK, Adinarayanan S. Comparison of intubation success and glottic visualization using King Vision and C-MAC videolaryngoscopes in patients with cervical spine injuries with cervical immobilization: A randomized clinical trial. *Surg Neurol Int*. 2017 Feb 6;8:19. doi: 10.4103/2152-7806.199560.
22. Gawlowski P, Iskryzcki L. Comparison of the KingVision video laryngoscope to the Macintosh laryngoscope for intubation of patient with cervical spine immobilization using extrication collars. *Am J Emerg Med*. 2017 Apr;35(4):657-658. doi: 10.1016/j.ajem.2016.12.064.
23. Passi Y, Sathyamoorthy M, Lerman J, Heard C, Marino M. Comparison of the laryngoscopy views with the size 1 Miller and Macintosh laryngoscope blades lifting the epiglottis or the base of the tongue in infants and children <2 yr of age. *Br J Anaesth*. 2014 Nov;113(5):869-74. doi: 10.1093/bja/aeu228.
24. Varghese E, Kundu R. Does the Miller blade truly provide a better laryngoscopic view and intubating conditions than the Macintosh blade in small children? *Paediatr Anaesth*. 2014 Aug;24(8):825-9. doi: 10.1111/pan.12394.
25. Hippard HK, Kalyani G, Olutoye OA, Mann DG, Watcha MF. A comparison of the Truview PCD and the GlideScope Cobalt AVL video-laryngoscopes to the Miller blade for successfully intubating manikins simulating normal and difficult pediatric airways. *Paediatr*

- Anaesth. 2016 Jun;26(6):613-20. doi: 10.1111/pan.12906.
26. Szarpak L, Kurowski A, Czyzewski L, Rodríguez-Nuñez A. Video rigid flexing laryngoscope (RIFL) vs Miller laryngoscope for tracheal intubation during pediatric resuscitation by paramedics: a simulation study. *Am J Emerg Med.* 2015 Aug;33(8):1019-24. doi: 10.1016/j.ajem.2015.04.054.
 27. Rodríguez-Núñez A, Moure-González J, Rodríguez-Blanco S, Oulego-Erroz I, Rodríguez-Rivas P, Cortiñas-Díaz J. Tracheal intubation of pediatric manikins during ongoing chest compressions. Does Glidescope® videolaryngoscope improve pediatric residents' performance? *Eur J Pediatr.* 2014 Oct;173(10):1387-90. doi: 10.1007/s00431-014-2329-z.
 28. Lee J, Kim JY, Kang SY, Kwak HJ, Lee D, Lee SY. Stylet angulation for routine endotracheal intubation with McGrath videolaryngoscope. *Medicine (Baltimore).* 2017 Feb;96(7):e6152. doi: 10.1097/MD.00000000000006152.
 29. Sakles JC, Kalin L. The effect of stylet choice on the success rate of intubation using the GlideScope video laryngoscope in the emergency department. *Acad Emerg Med.* 2012; 19(2):235-8. doi: 10.1111/j.1553-2712.2011.01271.x.
 30. Gombar S, Ahuja V. Malleable stylet in difficult intubation: a modified technique. *Br J Anaesth.* 2014 Aug;113(2):298-9. doi: 10.1093/bja/aeu246.
 31. Komasa N, Nishihara I, Minami T. Effects of stylet use during tracheal intubation on postoperative pharyngeal pain in anesthetized patients: A prospective randomized controlled trial. *J Clin Anesth.* 2017 May;38:68-70. doi: 10.1016/j.jclinane.2017.01.032.